

GOVERNMENT OF INDIA
CENTRAL INSTITUTE OF PSYCHIATRY
(General Section)

☎: 01-851-2451115(Exch) -2451113(O)

Fax: 01-851-2450823

E-mail: rch_cipra@sancharnet.in

Web: www.cipranchl.nic.in

Kanke, Ranchi- 834006, Jharkhand

No. 4910

Dated : 20.12.19


CORRIGENDUM

Reference Tender Notice No. 4793 Dated: 10.12.2019 issued vide this office and advertised in newspapers and also on website on 11.12.2019 is hereby amended to the following extent.

The following amendment made in the technical specification of item No 4 mentioned in List of Items to be purchased through Open Tender Notice-II (2019-20) i.e. Digital Radiography System with Single Detector for Whole Body Digital radiography.

Sl. No.	Technical Specification(Earlier)	Technical Specification may be read As (Amended)
1.	Technical Specification for Digital Radiography System with Single Detector for Whole Body Digital radiography	Technical Specification for Digital Radiography System with Dual Detector for Whole Body Digital radiography
2.	Tube Should be with anode heat storage capacity of 800 kHU or more	Tube Should be with anode heat storage capacity of 300 kHU or more

All the other terms & condition as mentioned in Tender Notice No. 4793 Dated: 10.12.2019 will be the same.


(Dr. D .Ram)
Director